

Ministry of Climate Change, Environment and Energy
Male', Republic of Maldives.

Answers to questions received with reference to Announcement Number (IUL)438-ENV/438/2023/540 dated 12th December 2023, Commissioning and Maintenance of a Continuous Ambient Air Quality Monitoring Station (CAAQMS)

21st December 2023

Please find below answers for clarification for the above bid.

- 1. Does the department already own any air quality monitoring stations? If yes, please can you provide details of supplier, brand, and model numbers.**
Yes, the Environmental Department under which this Project is managed has recently got one. But we do not know the details of the supplier
- 2. On page 62 – you do not state quantity required only the letter N has been displayed.**
It is required for the bidder to quantify
- 3. Page 63, item c.3 – 2.6kW aircon unit is not suitable for the enclosure sized 5 x 2.5 x 2.5**
Please state in the bid
- 4. Page 63, item c.4 – Please confirm that this is the correct size UPS.**
This is what is given by our Consultant. If you have recommendation. Please mention
- 5. Please advise the expected award date.**
Usually it takes about 6 weeks in the evaluation process.
- 6. May we know the location (Island) where the system and the Enclosure (House) will be installed?**
Male' City
- 7. As for the enclosure, is the size, 3 x 6 x 2.8 meters acceptable? (as this is a standard size for a galvanized container structure)**
We would require as per specified.
- 8. How Many workbench (No of people) will be stationed at the Enclosure.**
One Bench (about two personal).
- 9. Is a toilet required in the enclosure?**
No toilet is not required.

Procurement Section

Ministry of Climate Change, Environment and Energy
Green Building, Handhuvaree Hingun, Maafannu
Male', 20392, Republic of Maldives
Email: procurement@environment.gov.mv
Cc: cbit@environment.gov.mv
Website: www.environment.gov.mv